



PRODUCT CODE

NT-CE0540E Vicat Test Apparatus, EN complete set
NT-CE0540A Vicat Test Apparatus, ASTM complete Set
NT-CE0541E Vicat Mould, EN
NT-CE0542E Vicat initial needle, EN
NT-CE0543E Vicat final needle, EN
NT-CE0541A Vicat Mould, ASTM
NT-CE0542A Vicat needle, ASTM
NT-CE0544 Vicat transfer dish
NT-CE0545 Vicat Consistency Plunger, Ø10mm.



STANDARDS

EN 196-3, EN 480-2, AASHTO T131, ASTM C191

SPECIFICATIONS

The Vicat instrument is used to determine the setting time of cement paste and to measure its consistency.

The device consists of a metal frame, a graduated scale with index, a 300 g sliding probe, a Ø 10 mm consistency probe, an EN/ASTM Vicat mould, an EN/ASTM needle set and a glass base plate and a 110°C glass thermometer.

Accessories can be ordered seperatly if needed.

Accessories

NT-CE0541E Vicat Mould, EN
NT-CE0542E Vicat initial needle, EN
NT-CE0543E Vicat final needle, EN
NT-CE0541A Vicat Mould, ASTM
NT-CE0542A Vicat needle, ASTM
NT-CE0544 Vicat transfer dish
NT-CE0545 Vicat Consistency Plunger, Ø10mm.